

PROSPEKT - Clinical

- 309 *An Improved ON-line SPE-HPLC method for Cyclosporin A in whole blood samples - H. Herkdijk, J. Oui Koekstra - Spark Holland*
- 314 *Analysis of Serotonin in whole blood samples - A novel fully automated method to handle whole blood samples - C. Opper, W. Wesemann, G. Barka, H. Kerkdijk, G.Haak -LC-GC 12(9), 684-698 (1994)*
- 315 *On-line fully automated determination of clozapine desmethylclozapine in human serum by solid-phase extraction on exchangeable cartridges and liquid chromatography using methanol buffer mobile phase - O.V. Oleezen and B.Poulsen- J of Chrom.622(1993) 39-46*
- 316 *Automated analysis of urinary catecholamines by HPLC and on-line sample pretreatment - A.Pastoris, L.Cerutti, R. Sacco, L. De Vecchi and A. Sbaffi -J. of Chromatog. B 664 (1995), P.287-293*
- 333 *Improved High-Performance Liquid Chromatographic Method For The Determination Of Coenzyme Q10 In Plasma- G. Grossi, A. M. Bargossi, P.L. Fiorella, S.Piazzi, M.Battino, G.P. Bianchi- Journal Of Chromatography, 593 (1992) 217-226.*
- 517 *Direct Analysis Of Vitamin A In Serum By On-Line SPE-HPLC - H. Kerkdijk, O. Halmingh - Spark Oui Holland BV, The Netherlands.*
- 550 *Plasma Catecholamine levels in children - M.Candito, M. Albertini , P Chambon-Journal of Oui Chromatography, 617 (1993) 304-307.*
- 729 *Automated Determination of Vitamin A and E in Serum by One-line SPE-HPLC. by O Halmingh, Oui H.Kerkdijk, GJM. van Gils. SPARK Holland (March 1997)*
- 828 *A direct technique for the simultaneous determination of 10 drug candidates in plasma by liquid chromatography...by Debra A. McLoughlin, Timothy V. Olah & John D. Gilbert. Journal of Oui Pharmaceutical and Biomedical Analysis (1997), 15, P.1893-1901. ABSTRACT*
- 874 *Validation of an Automated Assay for Vitamin A and E in Serum by Hyphenated SPE-HPLC by O. Oui Halmingh, G. Haak and J.A. Ooms from Spark Holland B.V. October 1997*
- 1027 *On-Line Sample Preparation for Dermal Formulations-Method Development & Validation for ASM Oui 981 Cream & Liposomes, M. KRÄMER & A. MOYON, NOVARTIS Technical R&D Report N° 13/97*
- 1094 *Multiple automated clinical Assay's on a a single hyphenated SPE-HPLC-UV system, O. Oui HALMINGH, G.J.M. van GILS, J.A OOMS, SPARK HOLLAND*
- 1108 *Determination of ketorolac in human plasma by HPLC after automated on-line solid phase Non extraction, SOLA J., PRUNONOSA J., COLOM H., PERAIRE C. & OBACH R., J. Liq. Chromatogr. Relat. Technol (1996), 19 (1), PP.89-99*
- 1109 *Application of totally automated on-line sample clean up for prostanoid extraction and HPLC Non separation, HOTTER G., et al, Chromatographia, 36 (1993), PP.33-38.*
- 1118 *Liquid chromatography determination of the macrolide antibiotics roxithromycin and clarithromycin Non in plasma by automated phase extraction and electrochemical detection, HEDENMO M. & ERIKSSON BM., J. of Chromatog. A, 692 (1995), PP.161-166*

PROSPEKT - Clinical

- 1122 *Determination of the anti-platelet-activating factor BN-50727 and metabolites in human urine by solid-phase-extraction, Prunonosa, Sola, Peraire, Lavergne & Obach, J. of Chromatog. B, 677 (1996), PP.388-392*
Non
- 1123 *Automated high-performance liquid chromatographic assay for lesopitron, a novel anxiolytic, in human plasma using on*line solid-phase extraction, Garcia-Encina, Farran, Puig, Serafini & Martinez, J. of Chromatogr. B, 670 (1995), PP.103-110*
Non
- 1124 *Automated Solid-Phase Extraction and High-Performance Liquid Chromatographic Determination of Quinidine in Plasma, Brandsteterova A, et al, J. of Chromatogr. A, apr 8(1994), 665, PP101-104*
Non
- 1125 *Application of totally automated on-line sample clean up system for extraction and HPLC separation of peptide leukotrienes, RAMIS I, et al, J. of Pharmaceutical & Biomedical Analysis, Vol. 11, No 11/12, PP.1135-1139 (1993)*
Non
- 1170 *SPE-HPLC-FLU for Automated Profiling of Tryptophan Related Indoles in Plasma, I.P. Kema, G. Meiborg, O. Halmingh & J.A. Ooms, May 1999 -*
Oui
- 1171 *Second Generation Technology for High Throughput Hyphenated SPE-LC-MS and SPE-MS, Superfast Determination of Taxol in Serum - A. Schellen, B. Ooms, M. Van Gils, O. Halmingh, June 1999.*
Oui
- 1200 *Hyphenated On-Line SPE-HPLC for the Analysis of Dermal Cream Formulations by M. KRAMER, A. MOYON, O. HALMINGH & JA. OOMS, Spark Holland (May 1999)*
Oui
- 1254 *Generic SPE-LC-MS/MS Assay Protocol for fast Determination of Drugs in Biological Fluids, A. SCHELLEN (Diosynth), B.OOMS, M. van GILS & O. HALMINGH (Spark Holland), D. van de Lagemaat, W. Van Dongen, R. Vreeken & E. VERHEIJ (TNO Pharma), May 2000*
Oui
- 1287 *Fully automated determination of eserine N-oxide in human plasma using on-line SPE with LC coupled with electrospray ionization tandem MS, A. PRUVOST, I. RAGUENEAU, A. FERRY, P. JAILLON, JM GROGNET & H. BENECH, J. of Mass Spectro. (2000) 35, P625-633*
Oui